Importance of maintained atrio-ventricular synchrony in patients with pacemakers

J. Frielingsdorf*, A. E. Gerber and O. M. Hess*

*From the Cardiology Division, University Hospital, Zürich and the Cardiology Division, Triemli Hospital, Zürich, Switzerland

KEY WORDS: Dual chamber pacemaker, single chamber pacemaker, AV delay, work capacity, morbidity, mortality, quality of life.

The effect of atrial contraction on cardiac function is reviewed in patients with dual chamber and rate-responsive ventricular pacemakers. The question posed was is there any haemodynamic, clinical or prognostic advantage of AV synchrony in dual chamber pacemakers in comparison to rate-responsive ventricular pacemakers? Optimal AV delay in dual chamber pacing favours cardiac performance at rest, while during exercise the increase in heart rate rather than AV synchrony influences cardiac performance and working capacity. However, there is little information on the benefit of maintained AV synchrony in patients' daily activities. Patients with pacemakers which maintain AV synchrony seem to have less morbidity and mortality than patients with ventricular stimulation alone, and there are comparable rates of complication in carriers of single and dual chamber pacemakers, the former showing problems with the pacemaker syndrome and the latter with atrial sensing and pacemaker-induced tachycardias. The disadvantage of dual chamber pacemakers are higher costs and time-consuming controls.

Introduction

In comparison with fixed rate ventricular pacing (VVI), rate-responsive pacemakers (VVIR) may increase cardiac output and exercise capacity by a rise in heart rate^[1 5]. Dual chamber pacemakers (DDD) that sense the atrial wave, maintain the normal sequence of cardiac chamber activation (atrioventricular synchrony) and permit a chronotropic response to exercise in patients with normal sinoatrial function. During exercise, the additional atrial contribution to ventricular filling (atrial 'kick') in dual chamber pacing is supposed to improve haemodynamic function to a greater extent than in rate-responsive ventricular pacing alone. Any discussion about the importance of maintained atrioventricular (AV) synchrony has to answer the following: Is there any haemodynamic, clinical or prognostic advantage as regards AV synchrony in carriers of dual chamber pacemakers in comparison with rateresponsive ventricular pacemakers?

HAEMODYNAMIC ADVANTAGE OF MAINTAINED AV SYNCHRONY AT REST

Previous studies have shown that atrial contribution to ventricular filling in hearts with normal left ventricular function can cause a more than 20% increase in cardiac output^[6 13]. Particularly susceptible to this improvement are those patients with left ventricular

Submitted for publication on 8 November 1993, and in revised form 6 April 1994.

Correspondence J. Frielingsdorf, MD, Cardiology, University Hospital, 8091 Zürich, Switzerland.

hypertrophy^[14,15], those of advanced age^[16] or with left ventricular diastolic dysfunction^[17,18]. However, controversy exists about patients with impaired left ventricular systolic function. Several studies proved the value of AV synchrony in these patients^[18 20] but they did not evaluate the influence of left ventricular filling pressure on atrial contribution to cardiac output. Greenberg et al.^[21] and Myreng et al.^[22] found an inverse relationship between ventricular filling pressure and atrial contribution to cardiac output. Faerestrand and Ohm^[6], French et al. [23] and Dritsas et al. [24] found no benefit in relation to atrial systole in patients with left ventricular failure. Consequently, most evidence is demonstrated in hearts with normal ventricular filling pressures. Thus, at high passive filling pressures the ventricle is maximally dilated and operates at the extreme of the flattened Frank-Starling curve. Therefore, a further increase in preload by the atrial 'kick' does not augment stroke volume.

Irrespective of the underlying cardiac disease, in order to maximize cardiac output in patients with dual chamber pacemakers it is essential to have an appropriately timed atrial contraction prior to the onset of ventricular systole. The timing of atrial contraction is important in coordinating AV valve closure, since the major mechanism of such a closure is the increase in ventricular end-diastolic pressure^[18]. An inappropriately timed AV delay may cause mitral regurgitation and impaired diastolic filling^[25]. To avoid unnecessary ventricular pacing and early battery depletion, the AV delay is often lengthened to achieve an intrinsic ventricular contraction, but an impaired haemodynamic function has also been demonstrated in unpaced patients whose PR intervals are too long^[26,27].

Table 1 Optimal AV delay in patients at rest

| Reference, year | Parameters measured | Optimal AV delay (ms) | Remarks | | | | |
|--------------------------------------|------------------------|--------------------------|---|--|--|--|--|
| Nitsch ^[9] , 1984 | SV, CO | 100–150 | _ | | | | |
| Mehta ^[28] , 1989 | ĆO | 150 | | | | | |
| Iwase ^[29] , 1986 | LVIV | 150 | - | | | | |
| Wish ^[31] , 1987 | SV | 144 | 200 ms for long interatrial delays | | | | |
| Janosik ^[32] , 1989 | CO | 144 or 176 | 144 ms for VDD, 176 ms for DVI pacing | | | | |
| Frielingsdorf ^[33] , 1994 | EF | 150, 200 | _ | | | | |
| Sulke ^[34] , 1992 | Quality of life | 125, 175 RR | _ | | | | |
| Haskell ^[35] , 1986 | CO | 150, 200 | 150 ms for fast heart rates | | | | |
| Viden ^[36] , 1986 | EF, CI | 140-150 | 190-200 ms for patients with reduced EF | | | | |
| Dritsas ^[24] , 1993 | ŚV | 172, 142, 97 | 172 ms for diastolic dysfunction, 97 ms for systolic dysfunction, 142 ms for normal LV function | | | | |

AV=atrioventricular, CO=cardiac output, CI=cardiac index, DVI=AV sequential pacing, EF=ejection fraction, LV=left ventricular, LVIV=left ventricular inflow volume, RR=rate responsive AV delay, SV=stroke volume, VDD=atrial triggered ventricular pacing.

Even deviations in AV delay, by as little as 25 ms from an individual's optimal AV delay, resulted in a significant decrease in cardiac output^[26,28-32]. In a study by our group^[33] some patients showed a higher ejection fraction with single chamber ventricular pacing than with dual chamber pacing with an inappropriately timed AV delay (Table 1). Therefore, an inappropriately timed atrial systole as a result of a dual chamber pacemaker may be of little or no benefit as regards the haemodynamic effect than a ventricular-inhibited single chamber pacemaker^[9,13]. Little has been reported on the clinical relevance of optimizing the AV delay. Sulke *et al.*^[34] found that short AV delay settings (125 ms, 175 ms) were better tolerated than longer settings (250 ms) during out-of-hospital activity.

AV delays between 150 to 200 ms produced maximal left ventricular ejection fraction (Table 1). The optimal AV delay which makes atrial contraction coincide with late diastole and thus contribute most to left ventricular filling^[29] is determined by factors such as heart rate^[28,35], left ventricular function^[24,36–38], atrial size^[39] and the inter-individual variability of the intra-atrial conduction time^[31,32,40].

A shorter or longer intra-atrial conduction time corresponds to a shorter or longer AV delay^[31]. This is of special clinical relevance when pacing mode changes from dual chamber sequential (DVI) to atrial triggered ventricular pacing (VDD) because the left atrial to left ventricular sequence increases. To avoid a change in timing of left atrial contraction, with resulting decrease in stroke volume, the programmed AV delay must be shorter during atrial-triggered than during dual chamber sequential stimulation^[31,32,36]. A difference of 30–50 ms may be necessary to optimize cardiac output^[31,32,41] which emphasizes the importance of the AV delay hysteresis with altering pacing modes. This plays a particular role in the sick sinus syndrome which requires frequent atrial pacing.

Patients with hypertrophic cardiomyopathy or longstanding hypertension have abnormal ventricular diastolic function. The apparent increase in muscle stiffness and associated resistance to passive diastolic filling of the left ventricle underline the importance of an appropriately timed atrial systole in the preservation of ventricular filling in late diastole^[37,38]. Shefer et al.^[17] demonstrated striking haemodynamic differences between ventricular and AV sequential pacing. Loss of atrial synchrony was associated with a symptomatic fall in stroke volume. Dritsas et al. [24] observed that optimizing the AV delay is more important in patients with impaired left ventricular diastolic function than in patients with severely impaired systolic function. Therefore, efforts should be made to maintain AV synchrony in patients with abnormal diastolic function, e.g. hypertrophic cardiomyopathy, and, thus, an AV sequential pacemaker should be used.

HAEMODYNAMIC ADVANTAGE OF MAINTAINED AV SYNCHRONY DURING EXERCISE

Rate-adapted AV delay pacing

Whether a rate-adapted or a fixed short AV delay improves haemodynamic performance during exercise remains controversial (Table 2). A significant effect^[28,42,43] or no effect at all^[44-46] on haemodynamics or clinical parameters has been reported in patients with a shorter AV delay during exercise than at rest. The differing results might be explained by different study designs, especially when exercise capacity was only submaximal or maximal. Sheppard et al.^[47] investigated patients with complete AV block and found that at lower exercise levels cardiac haemodynamics did not differ when pacemakers were programmed to rate-adapted or fixed AV delays. At peak exercise, however, patients randomized to a rate-adapted AV delay showed better cardiac haemodynamics than patients with a fixed

Table 2 Benefit of an optimized AV delay during exercise

| Reference, year | Parameters measured | Method used | Remarks | | | | |
|--------------------------------|------------------------|-----------------------|---|--|--|--|--|
| With benefit | - | | | | | | |
| Mehta ^[28] , 1989 | СО | Doppler | AV delay of 75-100 ms preferred to AV delay of 150-200 ms | | | | |
| Leman ^[42] , 1985 | EF, SV | RNV | AV delay of 100 ms preferred to 150 ms | | | | |
| Ritter ^[43] , 1989 | CI, SVI | Thermodilution | Rate- adapted AV delay preferred to a fixed AV delay (150 ms, 200 ms) | | | | |
| Without benefit | | | , , , | | | | |
| Lau ^[44] , 1990 | CO | Doppler | Same CO for short (75 ms) and long 150 ms) AV delays | | | | |
| Haskell ^[45] , 1989 | Exercise capacity | 0 ₂ uptake | Same exercise capacity for short (66 ms) and long (168 ms) AV delays | | | | |
| Ryden ^[46] , 1988 | Exercise capacity | O ₂ uptake | Same exercise capacity for various AV delays (50, 100, 200, 250 ms) | | | | |

AV=atrioventricular, CI=cardiac index, CO=cardiac output, EF=ejection fraction. RNV=radionuclide ventriculography, SV=stroke volume, SVI=stroke volume index.

AV delay, although the duration of exercise was the same. The atrial contribution to ventricular filling is minimized or absent depending on the sensing of the P-wave, i.e. at higher heart rates induced by heavier exercise the P-wave coincides with the previous T-wave and, thus, eliminates the atrial 'kick'. Irrespective of the haemodynamic benefit, a rate-adapted AV delay seems to be preferable to a fixed AV delay which allows the upper tracking rate to increase with less limiting upper rate behaviour.

Dual chamber versus rate-adapted ventricular pacing

The same controversy exists whether dual chamber pacing is better than rate-responsive ventricular pacing. While some investigators^[8,44,48–50] demonstrated a slight but significant improvement in cardiac output during exercise due to AV synchrony, others[23,46,51-54] entirely attributed this improvement to the heart rate increase and not to AV synchrony (Table 3). This discrepancy is probably a matter of different pacemaker settings, posture, work loads, study populations and underlying heart disease. Lemke et al. [55] found in patients with isolated sick sinus node disease a more pronounced increase in oxygen consumption and work capacity during rate-responsive dual chamber pacing than during rate-responsive ventricular pacing. In patients with a high degree of AV block the differences between the pacing modes were not significant. Furthermore, the distinction was rarely made between patients with normal and impaired left ventricular function. French et al. demonstrated^[23] in patients with reduced left ventricular function, that during exercise physiological AV synchrony is less important that rate-responsiveness. During exercise the shortening of diastole unfavourably adds to high left ventricular filling pressures in the diseased heart; thus, the beneficial effects of AV synchrony may be even less. In a population with normal left ventricular function Ausubel et al.[56] showed that atrial synchronous pacing results in enhanced ventricular filling and spared contractile reserves when compared

with rate-responsive ventricular pacing, but heart rate, blood pressure and work load remained the same during exercise in the two pacing modes. In our own study^[50] none of the patients with normal or with impaired left ventricular function benefited clinically from a preserved AV synchrony; however, patients with normal left ventricular function and dual chamber pacing showed a higher oxygen consumption. Some of the discrepancies between the quoted studies may be accounted for by differences in exercise, which was either moderate or symptom-limited. Linde-Edelstam et al.[57] found that the relative contribution of atrial systole to ventricular filling is most important at rest, and that during exercise the relative contribution of atrial systole diminished concomitantly with a rise in mitral blood flow velocity. Thus, increasing kinetic energy of blood flow reduces atrial contribution to left ventricular filling.

QUALITY OF LIFE AND MAINTAINED AV SYNCHRONY

The effects of pacing mode on quality of life has been less studied. While Oldroyd et al.[58] and Bubien and Kay^[59] failed to find any difference in symptoms between dual chamber and rate-responsive ventricular pacing, others[60-64] demonstrated a better quality of life with dual chamber pacing. Linde-Edelstam et al.[61] explained the patients' preference for dual chamber pacing in comparison with rate-responsive ventricular pacing by a significant reduction in cardiovascular symptoms, improved self-perceived health, enhanced physical ability and psychological well-being. However, mean heart rate during the active hours of the day was higher with dual chamber pacing than with rateresponsive ventricular pacing. Menozzi et al. [62] investigated 14 patients with high degree AV block and found that the improvement in quality of life was reflected by the haemodynamic advantages of atrial synchronization. Five of these 14 patients did not tolerate rate-responsive ventricular pacing and were switched to dual chamber pacing because of intolerable symptoms, such as signs of

Table 3 Importance of pacing mode during exercise: VVIR versus AV synchronous (DDD, VDD) pacing

| Reference, year | Parameters measured | Method used | Remarks | | | |
|---------------------------------------|--------------------------|---------------------------------|--|--|--|--|
| AV with benefit | | | | | | |
| Karlöf ^[8] , 1975 | CO | CVC, Fick | - | | | |
| Lau ^[44] , 1990 | CO | Doppler | - | | | |
| Lemke ^[48] , 1990 | SV, ANP | CVC, Fick | ANP levels were equal | | | |
| Landzberg ^[49] , 1990 | Exercise capacity | Treadmill (time) | - | | | |
| Frielingsdorf ^[50] , 1993 | Exercise capacity | O ₂ uptake | patients with normal EF | | | |
| AV without benefit | | | | | | |
| French ^[23] , 1988 | Exercise capacity | 0 ₂ uptake | Patients with reduced EF | | | |
| Ryden ^[46] , 1988 | Exercise capacity | O ₂ uptake | - | | | |
| Frielingsdorf ^[50] , 1993 | Exercise capacity | O2 uptake | Patients with reduced EF | | | |
| Kristensson ^[51] , 1985 | SV, CO | Dye dilution | _ | | | |
| Pehrson ^[52] , 1983 | Exercise capacity | Bicycle (Watt) | - | | | |
| Fananapazir ^[53] , 1983 | Exercise capacity | Treadmill (time) | _ | | | |
| Linde-Edelstam ^[54] , 1992 | Exercise capacity SV, CO | O ₂ uptake CVC, Fick | Epinephrine and norepinephrine levels were equal | | | |

ANP=atrial natriuretic peptide, AV=atrioventricular, CO=cardiac output, CVC=central venous catheter, DDD=dual chamber pacing, SV=stroke volume, VDD=atrial triggered ventricular pacing, VVIR=rate-adapated ventricular pacing.

congestive heart failure and palpitations. These five patients were compared to nine patients who did not require the change to dual chamber pacing. The five patients were characterized by a lower incidence of organic heart disease and smaller echocardiographic chamber diameters. Sulke et al. [63] demonstrated that patients who preferred rate-responsive dual chamber pacing but felt that rate-responsive ventricular pacing was least acceptable had greater increases in stroke volume when paced in dual chamber mode than in the ventricular inhibited mode. Five out of 22 patients demanded early crossover because of intolerable symptoms, such as dyspnoea, dizziness and palpitations. Mitsuoka et al. [64] who investigated both patients with the sick sinus syndrome and retrograde atrioventricular conduction and those with high degree AV block, demonstrated beneficial effects on symptoms of dual chamber pacing compared with ventricular pacing.

The weakness of most studies which deal with quality of life is that AV delay is not appropriately timed and the programming of rate-responsive parameters is not optimized. A programmed activity sensor that is too sensitive with inappropriately high heart rates has been demonstrated to be poorly tolerated by patients^[65].

Sulke et al. [66] reported that most patients who were satisfied with long-term VVI pacing benefited from the upgrading of ventricular to dual chamber pacing; this suggests there is a 'subclinical' pacemaker syndrome. The disadvantage of ventricular pacing is related to the absence of — or random occurrence of — atrial contri-

bution to ventricular filling; thus, stroke volume is especially decreased at rest. A further unfavourable factor is the presence of retrograde atrial contraction (VA conduction) which occurs simultaneously or just after the paced ventricular contraction. The lack of coordination between atrial and ventricular contraction may provoke palpitations and hypotension with symptoms of dizziness which have been described as the 'pacemaker syndrome' [67]. The atrial contractions against closed AV valves, which induce retrograde atrial emptying and atrial distension, may initiate unfavourable peripheral autonomic reflexes [68 70] or blunt the baroreceptor reflex; cardiac output sometimes decreases to levels lower than during unpaced sinus bradycardia. Patients with the sick sinus syndrome are particularly prone to develop the pacemaker syndrome because they often have preserved AV conduction. Up to 90% of such patients may have a VA conduction^[71], whereas the cumulative incidence of the pacemaker syndrome has been reported as high as 7% and as low as $0.5\%^{[72,73]}$. The haemodynamic compromise observed with VA conduction appears disproportionately greater than would be expected if it were the mere consequence of the loss of atrial transport. VA conduction may cause atrial contraction to coincide with a closed AV valve and produce pronounced atrial distention. Nishimura et al.[74] demonstrated that nearly all patients with symptomatic ventricular-inhibited pacing-induced hypotension had intact VA conduction. Another entity is the 'AAIR pacemaker syndrome' [75,76]. Patients chronically paced

Table 4 Atrial fibrillation, chronic heart failure and mortality in ventricular (VVI) versus atrial based pacing (AAI, AV synchronous) in patients with the sick sinus node syndrome or AV block

| Refer e nce, y c ar | Follow-up (months) | Patients (n) | | AF (%) | | CHF (%) | | Mortality (%) | |
|--|-----------------------|--------------|--------|--------|---------------------------------------|---------|--------|---------------|--------|
| | | VVI | AAI/AV | VVI | AAI/AV | VVI | AAI/AV | VVI | AAI/AV |
| Sick sinus node | | | | | · · · · · · · · · · · · · · · · · · · | | _ | | |
| Stangl ^[77] , 1990 | 53 | 112 | 110 | 19 | 6 | _ | _ | 31 | 19 |
| Feuer ^[78] , 1989 | 44 | 70 | 61 | 25 | 11 | _ | | _ | _ |
| Rosenqvist ^[80] , 1986 | 24 | 79 | 89 | 30 | 4 | 23 | 7 | 10 | 6 |
| Markewitz ^[81] , 1986 | 32 | 87 | 136 | 30 | 7 | _ | _ | _ | _ |
| Santini ^[83] , 1990 | 54 | 125 | 214 | 47 | 7 | _ | _ | 30 | 14 |
| Rosenqvist ^[84] , 1988 | 48 | 79 | 89 | 47 | 7 | 37 | 15 | 23 | 8 |
| Sasaki ^[85] , 1991 | 62 | 34 | 41 | 44 | 17 | 21 | 2 | 24 | 0 |
| Zanini ^[84] , 1989 | 40 | 57 | 53 | 18 | 4 | 5 | 2 | 18 | 9 |
| Sasaki ^[87] , 1988 | 35 | 25 | 24 | 36 | 0 | 28 | 4 | 24 | 13 |
| Andersen ^[88] , 1993 | • | 115 | 110 | 19 | 10 | | | 13 | 11 |
| Albert ^[93] , 1987 | 60 | 79 | 49 | _ | | _ | | 26 | 16 |
| Mean** | | | | 28 | 7 | 24 | 6 | 24 | 12 |
| AV block | | | | | | | | | |
| Alpert ^[94] , 1986 | 60 | 132 | 48 | _ | _ | _ | _ | 34 | 15 |
| Linde ^[95] , 1992 | 65 | 74 | 74 | _ | _ | _ | _ | 36 | 32 |
| Mean | | | | | | | | 35 | 25 |

AAI=atrial pacing, AV=atrioventricular synchronous pacing, AF=atrial fibrillation, atrial flutter, CHF=congestive heart failure, *no data; **study no. 80 was excluded because it was amalgamated with no. 84

in the rate-responsive atrial mode may have a paradoxically increased stimulus-R interval, producing P waves occurring immediately after, or even within, the R wave of the preceding cycle which nullifies the haemodynamic contribution of the atrial systole. This may result in a limitation of the exercise capacity.

MORBIDITY AND MORTALITY WITH AND WITHOUT MAINTAINED AV SYNCHRONY

Morbidity: Atrial fibrillation, thromboembolic complications and chronic heart failure (Table 4, Fig. 1)

In patients with a sick sinus node disease the incidence of a new onset of atrial fibrillation, with its associated




Figure 1 Percentage of chronic atrial fibrillation/flutter (AF), chronic heart failure (CHF) and mortality with ventricular (VVI, □) and atrial based (AAI/AV, ■) pacing, reported in patients with the sick sinus node disease (average of Table 4). The percentage of patients is given on top of each bar, n=total number of patients.

thromboembolic complications, is higher with ventricular stimulation (VVI) than with a stimulation in which AV synchrony is preserved (atrial based pacing: AAI, VDD, DDD). The review of several studies[77-88] reveals an incidence of atrial fibrillation of 18-47% among patients with ventricular stimulation and 0-17% of all patients with AV synchronous stimulation. A literature review by Sutton and Kenny^[89] demonstrated over a period of 31 months a 12.3% incidence of thromboembolic complications in patients with ventricular stimulation (n = 347) in comparison with 1.6% (P < 0.001) with atrial stimulation (n=321). The incidence of chronic heart failure for both ventricular and AV synchronous pacing resembles the corresponding incidence of atrial fibrillation^[84–87] for these pacing modes. There are no data available on patients with high degree AV block and pacing modality as regards atrial fibrillation or chronic heart failure.

Mortality (Table 4, Fig. 1)

The 5-year survival rate of patients with a sick sinus node disease is, on the one hand, comparable with the mortality of patients with high degree AV block, and on the other with mortality of the non-paced population^[90-92]. The prognosis for survival is related to the underlying heart disease and left ventricular function. The influence of pacing modality on survival has been addressed, in particular, in patients with sinus node disease and it is the object of only a few studies^[77,83 88,93]. Rosenqvist et al.^[84] reported a mortality of 23% in the population with ventricular stimulation in comparison with the significantly lower mortality of 8% in the atrial paced population. Alpert et al.^[93] demonstrated a lower survival rate in patients with chronic heart failure — but not in patients without heart

failure — and ventricular stimulation than in patients with dual chamber stimulation. In patients with high degree AV block, AV synchronous pacing was preferable to ventricular pacing only when congestive heart failure was present^[94,95].

The quoted studies demonstrate the superiority of AV synchronous pacing to ventricular pacing. However, some critical remarks are necessary: the conclusions of these studies only apply to paced patients with a sick sinus node disease. Patients with high degree AV block were not sufficiently investigated. Apart from two exceptions^[85,88], all data originate from retrospective rather than randomized prospective trials. In some studies the study population was small and patients with fixed-rate ventricular stimulation served as controls. There are only limited data available which compare rate-responsive ventricular with AV synchronous stimulation. Van Erckelens et al.^[82] showed a similar incidence of atrial fibrillation with rate-responsive ventricular and dual chamber pacing.

In patients with the sick sinus node disease and without AV conduction abnormalities, atrial inhibited pacing may be an alternative to dual chamber pacing. A study by Elshot et al. [96] with a follow-up of 12 years revealed an annual incidence of permanent atrial fibrillation of 1% which is comparable to the results of other recent investigations^[97,98]. In the same study^[96] the annual incidence of chronic atrial fibrillation in atrial pacing was reported from several other studies to range between 0.4 and 3.3%, compared to 4.3 and 11.7% in ventricular pacing; the reported annual mortality rate was between 0.5 and 4.8% in atrial pacing compared to 3.5 and 9.3% in ventricular pacing. Chronic atrial fibrillation mainly develops in patients with the brady-tachy syndrome^[96] or in patients with a long history of episodes of pre-implant paroxysmal atrial fibrillation lasting more than 1 h^[99]. Atrial inhibited pacing in the sick sinus node disease may be limited by progression of high degree AV block, but two reviews of 28^[89] and 27 studies^[100], respectively, showed a low annual incidence of AV conduction disturbances ranging between 0.7 and 3%.

Atrial inhibited pacing might even be superior to dual chamber pacing, partly due to the prevention of bradycardia-induced inhomogeneity of atrial refractoriness and partly to improved haemodynamics. There is evidence that atrial pacing may prevent paroxysmal atrial fibrillation^[101,102]. Haemodynamic characteristics are better with atrial-inhibited pacing compared to dual chamber pacing^[103-105] which is partly explained by intrinsic AV conduction^[106] and synchronous ventricular contraction^[107,108]. To our knowledge, there are no prospective morbidity and mortality studies comparing atrial inhibited pacing to dual chamber pacing.

COMPLICATIONS AND COSTS

Complications

The implantation of a dual chamber pacemaker is not associated with a higher rate of complications than that

of a single chamber pacemaker. Whereas the most frequent complication of ventricular stimulation is the pacemaker syndrome, the dual chamber pacemaker is prone to malfunction of atrial sensing and pacemakerinduced tachycardia. Mueller et al. [109] prospectively analysed 258 patients with single chamber (VVI) and 75 patients with dual chamber pacemakers over a period of 3 years. There was no significant difference in relation to reintervention (5% of single chamber and 5.3% of dual chamber pacemakers), lead displacement with reoperation (1% vs 1·3%), infection (0·77% vs 1·33%), muscular stimulation (3.5% vs 4%) and urgent reprogramming (9% vs 8%). Goldman et al.[110] showed a higher perioperative complication rate in patients with dual chamber pacemakers (6.8%) than in patients with single chamber pacemakers (2.6%). In the study of Parsonnet et al.[111] there was a substantially larger incidence of complications when the implanters performed less than 12 implantations per year, this was particularly true for the incidence of lead-related complications.

Costs

Over a 12-year period Eagle et al.[112] compared the incremental cost of single and dual chamber pacemakers. At implantation, the single chamber device costed US \$2503 less than the dual chamber device (\$6924 compared with \$9427) and after 12 years the cost difference was \$5167 (\$11 339 compared with \$16 506). Thus, the difference in cost between single and dual chamber pacemakers is not simply a reflection of the initial implantation costs. The cumulative difference in costs continues to increase with time, and the major contribution was the more frequent failure rate of dual chamber generators. In this 12-year frame, a recipient of a dual chamber pacemaker would anticipate 0.74 generator replacements compared to 0.46 for single chamber pacemaker recipients. The differences would have been even greater if patients had survived longer. The quoted study was published in 1986; thus, as a result of improved survival in dual chamber generator batteries a third of the eventual cost difference could be eliminated. However, one of the major contributors to postimplantation costs, which remained unchanged over time, was the more frequent follow-up schedule for patients with dual chamber devices[112]. In the case of single chamber ventricular devices, the pacemaker syndrome and its resulting need for pacemaker revision to a dual chamber system accounted for just a small fraction of total cost[112]. A recent British study[113], which covered Northern England, with a population of three million inhabitants, showed considerable changes in regional pacing practice. The adoption of the optimal pacing practice recommended by the British Pacing and Electrophysiology Groups, e.g. dual chamber instead of single chamber pacing, would result in a substantial increase in expenditure. According to these recommendations the annual pacing budget for pacing hardware would have increased from £333 535 to £647 163 (increase of 94%) for the use of the optimal pacing mode. But these studies have not taken into account lower

morbidity and mortality associated with dual chamber pacing, resulting in a considerable health benefit, thus, reducing rehospitalizations and costs. There might be a substantial reduction in costs if atrial pacemakers were more widely used than dual chamber systems in patients with a sick sinus node disease. Costs could also be saved by changing a single chamber atrial pacemaker to either a ventricular pacemaker (because of chronic atrial fibrillation) or to a dual chamber pacemaker (because of progression of AV conductance disturbances).

Conclusions

The effect of atrial contraction on cardiac function is complex and depends on different interacting factors. such as left ventricular systolic and diastolic function, intra-atrial conduction time as well as atrial size and contraction. It seems reasonable to conclude that dual chamber pacing is superior to ventricular pacing, as regards haemodynamics at rest and quality of life. The data on optimal pacing during exercise conditions would require more homogeneous study designs for the determination of the importance of maintained AV synchrony, such as work loads, optimized AV delays and study populations. Furthermore, it remains to be determined whether other groups such as older patients, less active patients or those with diastolic dysfunction may especially benefit from AV synchrony. Patients with pacemakers which maintain AV synchrony seem to have a lower morbidity and mortality than patients with single chamber ventricular stimulation, but larger prospective randomized trials are needed to prove or disprove this hypothesis.

Atrial inhibited pacing is an under-used mode of cardiac stimulation but still remains the best pacing mode in many patients with the sick sinus node disease, whereas dual chamber pacing is the preferred mode in patients with high degree AV block. Today, single chamber ventricular pacing is restricted to patients with chronic atrial fibrillation and brady-arrhythmias.

References

- [1] Faerestrand S, Breivik K, Ohm OJ. Assessment of the work capacity and relationship between rate response and exercise tolerance associated with activity-sensing rate-response pacing. Pace 1987; 10: 1277-90.
- [2] Humen DP, Kostuk WJ, Klein GJ. Activity-sensing, rateresponsive pacing: Improvement in myocardial performance with exercise. Pace 1985; 8: 52-9.
- [3] Lipkin DP, Buller N, Frenneaux M et al. Randomized crossover trial of rate responsive Activitrax and conventional fixed rate ventricular pacing. Br Heart J 1987; 58: 613-6.
- [4] Benditt DG, Mianulli M, Fetter J et al. Single- chamber cardiac pacing with activity initiated chronotropic response: Evaluation by cardiopulmonary exercise testing. Circulation 1987; 75: 184-91.
- [5] Tilbury RT, Vlietstra RE, Hayes DL. Activity-mode cardiac pacing: Clinical considerations. Mayo Clin Proc 1987; 62: 607-11.
- [6] Faerestrand S, Ohm OJ. A time-related study of the hemodynamic benefit of atrioventricular synchronous pacing evaluated by Doppler echocardiography. Pace 1985; 8: 838-48.

- [7] Kruse I, Arnman K, Conradson TB, Rydén L. A comparison of the acute and long-term hemodynamic effects of ventricular inhibited and atrial synchronous ventricular inhibited pacing. Circulation 1982; 65: 846-55.
- [8] Karlöf I. Haemodynamic effect of atrial triggered versus fixed rate pacing at rest and during exercise in complete heart block. Acta Med Scand 1975; 197: 195-206.
- [9] Nitsch J, Seiderer M, Büll U, Büderitz B. Evaluation of left ventricular performance by radionuclide ventriculography in patients with atrioventricular versus ventricular demand pacemakers. Am Heart J 1984; 107: 906-11.
- [10] McMeekin JD, Lautner D, Spencer H, Gulamhusein SS. Importance of heart rate response during exercise in patients using atrioventricular synchronous and ventricular pacemakers. Pace 1990; 13: 59-68.
- [11] Curtis JJ, Maloney JD, Barnhorst DA, Pluth JR, Hartzler GO, Wallace RB. A critical look at temporary ventricular pacing following cardiac surgery. Surgery 1977; 82: 888-93.
- [12] Stewart WJ, Dicola VC, Harthrone JW, Gillam LD, Weyman AE. Doppler ultrasound measurement of cardiac output in patients with physiologic pacemakers. Am J Cardiol 1984; 54: 308-12.
- [13] Curtis JJ, Madigan NP, Whiting RB et al. Clinical experience with permanent atrioventricular sequential pacing. Ann Thorac Surg 1981; 32: 179–85.
- [14] Bonow RO, Frederick TM, Bacharach SL et al. Atrial systole and left ventricular filling in hypertrophic cardiomyopathy: Effect of Verapamil. Am J Cardiol 1983; 51: 1386-91.
- [15] Johnson AD, Daily PO. Hypertrophic subaortic stenosis complicated by high degree heart block: Successful treatment with an atrial synchronous ventricular pacemaker. Chest 1975; 67: 491-4.
- [16] Miyatake K, Okamoto M, Kinoshita N et al Augmentation of atrial contribution to left ventricular inflow with aging as assessed by intracardiac Doppler flowmetry. Am J Cardiol 1984; 53: 586-9.
- [17] Shefer A, Rozenman Y, David YB, Flugelman MY, Gotsman MS, Lewis BS. Left ventricular function during physiological cardiac pacing: Relation to rate, pacing mode, and underlying myocardial disease. Pace 1987; 10: 315-25.
- [18] Rahimtoola SH, Ehsani A, Sinno MZ, Loeb HS, Rosen KM, Gunnar RM. Left atrial transport function in myocardial infarction. Am J Med 1975; 59: 686-94.
- [19] Reiter MJ, Hindman MC. Hemodynamic effects of acute atrioventricular sequential pacing in patients with left ventricular dysfunction Am J Cardiol 1982; 49: 687–92.
- [20] Benchimol A, Ellis JG, Dimond EG. Hemodynamic consequences of atrial and ventricular pacing in patients with normal and abnormal hearts. Am J Med 1965; 39: 911–22.
- [21] Greenberg B, Chatterjee K, Parmley WW, Werner JA, Holly AN. The influence of left ventricular filling pressure on atrial contribution to cardiac output. Am Heart J 1979; 98: 742-51.
- [22] Myreng Y, Smiseth OA, Risoe C. Left ventricular filling at elevated diastolic pressures: Relationship between transmitral Doppler flow velocities and atrial contribution. Am Heart J 1990; 119: 620-6.
- [23] French WJ, Haskell RJ, Wesley GW, Florio J. Physiological benefits of a pacemaker with dual chamber pacing at low heart rates and single chamber rate responsive pacing during exercise. Pace 1988; 11 (II): 1840-5.
- [24] Dritsas A, Joshi J, Webb S, Koukoulas A, Oakley CM. Nihoyannopoulos P. Optimal atrioventricular interval during dual-chamber pacing: relation to underlying ventricular function. Eur Heart J 1993; 14(Suppl): 252 (Abstr).
- [25] Mitchell JH, Gilmore JP, Sarnoff SJ. The transport function of the atrium. Am J Cardiol 1962; 9: 237.
- [26] Hartzler GO, Maloney JD, Curtis JJ, Barnhorst DA. Hemodyanamic benefits of atrioventricular sequential pacing after cardiac surgery. Am J Cardiol 1977; 40: 232-6.
- [27] Benchimol A, Duenas A, Liggett MS, Dimond EG. Contribution of atrial systole to the cardiac function at a fixed and at a variable ventricular rate. Am J Cardiol 1965; 16: 11-21.

- [28] Mehta D, Gilmour S, Ward DE, Camm AJ. Optimal atrioventricular delay at rest and during exercise in patients with dual chamber pacemakers. A non-invasive assessment by continuous wave Doppler. Br Heart J 1989; 61: 161-6
- [29] Iwase M, Sotobata I, Yokota M et al. Evaluation by pulsed Doppler echocardiography of the atrial contribution to left ventricular filling in patients with DDD pacemakers. Am J Cardiol 1986; 58: 104-9.
- [30] Coskey RL, Feit TS, Plaia R, Zicari T. AV pacing and LV performance. Pace 1983; 6: 631-40.
- [31] Wish M, Fletcher RD, Gottdiener JS, Cohen AI. Importance of left atrial timing in the programming of dual-chamber pacemakers. Am J Cardiol 1987; 60: 566-71
- [32] Janosik DL, Pearson AC, Buckingham TA, Labovitz AJ, Redd RM. The hemodynamic benefit of differential atrioventricular delay intervals for sensed and paced atrial events during physiologic pacing. J Am Coll Cardiol 1989; 14: 499-507.
- [33] Frielingsdorf J, Gerber AE, Dür P, Vuilliomenet A, Bertel O. Importance of an individually programmed atrioventricular delay at rest and on work capacity in patients with dual chamber pacemakers. Pace 1994; 17: 37-45.
- [34] Sulke AN, Chambers JB, Sowton E. The effect of atrioventricular delay programming in patients with DDDR pacemakers. Eur Heart J 1992; 13: 464-72.
- [35] Haskell RJ, French WJ. Optimum AV interval in dual chamber pacemakers. Pace 1986; 9: 670-5.
- [36] Videen JS, Huang SK, Bazgan ID, Mechling E, Patton DD Hemodynamic comparison of ventricular pacing, atrioventricular sequential pacing and atrial synchronous ventricular pacing using radionuclide ventriculography Am J Cardiol 1986; 57: 1305–8.
- [37] Fananapazir L, Connon RO, Tripodi D, Panza JA. Impact of dual chamber permanent pacing in patients with obstructive hypertrophic cardiomyopathy with symptoms refractory to verapamil and beta-adrenergic blocker therapy. Circulation 1992; 85: 2149-61.
- [38] McDonald K, McWilliams, Maurer B. Improved functional capacity in patients treated with dual chamber pacing for hypertrophic cardiomyopathy. Pace 1987; 107: 714.
- [39] Labovitz AJ, Williams GA, Redd RM, Kennedy HL. Noninvasive assessment of pacemaker hemodynamics by Doppler echocardiography: Importance of left atrial size. J Am Coll Cardiol 1985; 6: 196-200.
- [40] Bachmann G. The inter-auricular time interval. Am J Physiolog 1916; 41: 309-20.
- [41] Alt E, von Bibra H, Blomer H. Different beneficial AV intervals with DDD pacing after sensed or paced atrial events. J Electrophysiol 1987; 1: 250-6.
- [42] Leman RB, Kratz JM. Radionuclide evaluation of dual chamber pacing: Comparison between variable AV intervals and ventricular pacing. Pace 1985; 8: 408-14.
- [43] Ritter P, Daubert C, Mabo P, Descaves C, Gouffault J. Haemodynamic benefit of a rate-adapted AV delay in dual chamber pacing. Eur Heart J 1989; 10: 637-46.
- [44] Lau CP, Wong CK, Leung WH, Liu WX. Superior cardiac hemodynamics of atrioventricular synchrony over rate responsive pacing at submaximal exercise: Observations in activity sensing DDDR pacemakers. Pace 1990; 13 (II): 1832-7.
- [45] Haskell R, French WJ. Physiological importance of different atrioventricular intervals to improved exercise performance in patients with dual chamber pacemakers. Br Heart J 1989; 61: 46-51.
- [46] Ryden L, Karlsson O, Kristensson BE. The importance of different atrioventricular intervals for exercise capacity. Pace 1988: 11: 1051-62.
- [47] Sheppard RC, Ren JF, Ross J, McAllister M, Chandrasekaran, Kutalek SP. Doppler echocardiographic assessment of the hemodynamic benefits of rate adaptive AV delay during exercise in paced patients with complete heart block. Pace 1993; 16: 2157-67.

- [48] Lemke B, Gude J, v.Dryander S, Barmeyer J, Braun BE, Krieg M. Einfluss von AV-Synchronisation und Frequenzsteigerung auf die Hämodynamik und das atrial natriuretische Peptid bei Patienten mit totalem AV-Block. Z Kardiol 1990; 79: 547-56.
- [49] Landzberg JS, Franklin JO, Mahawar SK et al. Benefits of physiologic atrioventricular synchronization for pacing with an exercise rate response. Am J Cardiol 1990; 66: 193-7.
- [50] Frielingsdorf J, Gerber A, Dür P, Vuilliomenet A, Amann FW, Bertel O. Rate responsive single chamber (VVIR) versus dual chamber pacing (DDD) and work capacity: Role of left ventricular function Eur Heart J 1993; 14 (Abstr Suppl): 127
- [51] Kristensson BE, Arnman K, Rydeén L. The hemodynamic importance of atrioventricular synchrony and rate increase at rest and during exercise. Eur Heart J 1985; 6: 773-8.
- [52] Pehrsson SK. Influence of heart rate and atrioventricular synchronization on maximal work tolerance in patients treated with artificial pacemakers. Acta Med Scand 1983; 214: 311-5.
- [53] Fananapazir L, Bennett DH, Monks P. Atrial synchronized ventricular pacing: Contribution of the chronotropic response to improved exercise performance. Pace 1983; 6 (I): 601-8.
- [54] Linde-Edelstam C, Hjemdahl P, Pehrsson SK, Aström H, Nordlander R. Is DDD pacing superior to VVI,R? A study on cardiac sympathetic nerve activity and myocardial oxygen consumption at rest and during exercise. Pace 1992; 15 (I): 425-34.
- [55] Lemke B, Dryander SV, Jäger D, Machraoui A, MacCarter D, Barmeyer J. Aerobic capacity in rate modulated pacing. Pace 1992; 15 (I): 1914-8.
- [56] Ausubel K, Steingart RM, Shimshi M, Klementowicz P, Furman S. Maintenance of exercise stroke volume during ventricular versus atrial synchronous pacing: Role of contractility. Circulation 1985; 72: 1037-43.
- [57] Linde-Edelstam C, Juhlin-Dannfelt A, Nordlander R, Pehrsson SK. The hemodynamic importance of atrial systole: A function of the kinetic energy of blood flow? Pace 1992; 15 (I): 1740-9.
- [58] Oldroyd KG, Rae AP, Carter R et al. Double blind crossover comparison of the effects of dual chamber pacing (DDD) and ventricular rate adaptive (VVIR) pacing on neuroendocrine variables, exercise performance, and symptoms in complete heart block. Br Heart J 1991; 65: 188-93.
- [59] Bubien RS, Kay GN. A randomized comparison of quality of life and exercise capacity with DDD and VVIR pacing modes. Pace 1990; 13: 524 (Abstr).
- [60] Perrins EJ, Morley CA, Chan SL, Sutton R. Randomized controlled trial of physiological and ventricular pacing. Br Heart J 1983; 50: 112-7.
- [61] Linde-Edelstam C, Nordlander R, Undén AL, Orth-Gomér K, Rydén L. Quality-of-life in patients treated with atrioventricular synchronous pacing compared to rate modulated ventricular pacing: A long-term, double-blind, crossover study. Pace 1992; 15 (I): 1467-76.
- [62] Menozzi C, Brignole M, Moracchini PV et al. Intrapatient comparison between chronic VVIR and DDD pacing in patients affected by high degree AV block without heart failure. Pace 1990; 13 (II): 1816-22.
- [63] Sulke N, Chambers J, Dritsas A, Sowton E. A randomized double- blind crossover comparison of four rate-responsive pacing modes. J Am Coll Cardiol 1991; 17: 696-706.
- [64] Mitsuoka T, Kenny RA, Yeung TA, Chan SL, Perrins JE, Sutton R. Benefits of dual chamber pacing in sick sinus syndrome. Br Heart J 1988; 60: 338-47.
- [65] Sulke AN, Dritsas A, Chambers J et al. Is accurate rate response programming necessary? Pace 1990; 13: 1031-44.
- [66] Sulke N, Dritsas A, Bostock J, Wells A, Morris R, Sowton E. 'Subclinical' pacemaker syndrome: A randomized study of symptom free patients with ventricular demand (VVI) pacemakers upgraded to dual chamber devices. Br Heart J 1992; 67: 57-64.

- [67] Mitsui T, Hori M, Suma K, Wanibuchi Y, Saigusa M. The 'pacemaker syndrome'. In: Jacobs JE. Proceedings of the Eighth Annual International Conference on Medical and Biological Engineering. Chicago: Association for the Advancement of Medical Instrumentation. 1969: 29-33.
- [68] Alicandri C, Found FM, Tarazi RC et al. Three cases of hypotension and syncope with ventricular pacing. Possible role of atrial reflexes. Am J Cardiol 1978; 42: 137-42.
- [69] Erlebacher JA, Danner RL, Stelzer PE. Hypotension with ventricular pacing: An atrial vasodepressor reflex in human beings. J Am Coll Cardiol 1984; 4: 550-5.
- [70] Ausubel K, Boal BH, Furman S. Pacemaker syndrome: definition and evaluation. Card Clinic 1985; 3: 587-94.
- [71] Morley C, Sutton R, Perrins J et al. Importance of retrograde atrioventricular conduction in physiological cardiac pacing. Pace 1981; 3: A-60.
- [72] Parker MM, Lemberg L. Pacemaker update 1984: The pacemaker syndrome. Heart Lung 1984; 13: 448-50.
- [73] Cohen SI, Frank HA. Preservation of active atrial transport: An important clinical consideration in cardiac pacing. Chest 1982; 81: 51-4.
- [74] Nishimura RA, Gersh BJ, Vlietstra RE, Osborn MJ, Ilstrup DM, Holmes DR, Jr. Hemodynamic and symptomatic consequences of ventricular pacing. Pace 1982; 5: 903–10.
- [75] Den Dulk K, Lindemans FW, Brugada P et al. Pacemaker syndrome with AAI rate variable pacing. Importance of atrio-ventricular conduction properties, medication and pacemaker programmability. Pace 1988; 11: 1226-33.
- [76] Mabo P, Pouillot C, Kermarrec A, Lelong B, Lebreton H, Daubert C. Lack of physiological adaption of the atrioventricular interval to heart rate in patients chronically paced in the AAIR mode Pace 1991; 14: 2133-42.
- [77] Stangl K, Seitz K, Wirtzfeld A et al. Differences between atrial single chamber pacing (AAI) and ventricular single chamber pacing (VVI) with respect to prognosis and antiarrhythmic effect in patients with sick sinus syndrome Pace 1990; 13 (II): 2080-5.
- [78] Feuer J, Shandling A, Messenger J, Castellanet CD, Thomas LA. Influence of cardiac pacing mode on the long-term development of atrial fibrillation. Am J Cardiol 1989, 64. 1376-9
- [79] Langenfeld H, Grimm W, Maisch B, Kochsiek K. Atrial fibrillation and embolic complications in paced patients. Pace 1988; 11 (II): 1667-72
- [80] Rosenqvist M, Brandt J, Schuller H. Atrial versus ventricular pacing in sinus node disease. A treatment comparison study Am Heart J 1986; 111: 292-7.
- [81] Markewitz A, Schad N, Hemmer W, Bernheim C, Ciavolella M, Weinhold C. What is the most appropriate stimulation mode in patients with sinus node dysfunction? Pace 1986; 9 (11): 1115-20.
- [82] Van Erckelens F, Sigmund M, Lambertz H, Kreis A, Hanrath P. Atrial fibrillation in rate responsive-, DDD- and VVI-pacing: Incidence, time of onset, influencing factors. Pace 1990, 13: 1194 (Abstr).
- [83] Santini M, Alexidou G, Ansalone G, Cacciatore G, Cini R, Turitto G. Relation of prognosis in sick sinus syndrome to age, conduction defects and modes of permanent cardiac pacing. Am J Cardiol 1990; 65: 729-35.
- [84] Rosenqvist M, Brandt J, Schuller H. Long-term pacing in sinus node disease: Effects of stimulation mode on cardiovascular morbidity. Am Heart J 1988; 116: 16-22.
- [85] Sasaki Y, Furihata A, Suyama K et al. Comparison between ventricular inhibited pacing and physiologic pacing in sick sinus syndrome. Am J Cardiol 1991; 67: 771-4.
- [86] Zanini R, Facchinetti A, Gallo G et al. Morbidity and mortality of patients with sinus node disease: Comparative effects of atrial and ventricular pacing. Pace 1990; 13 (II): 2076 9.
- [87] Sasaki Y, Shimotori M, Akahane K et al. Long-term follow-up of patients with sick sinus syndrome: A compari-

- son of clinical aspects among unpaced, ventricular inhibited paced and physiologically paced groups. Pace 1988; 11 (I): 1575-83.
- [88] Andersen HR, Thuesen L, Bagger JP, Vesterlund T, Bloch Thomsen PE. Atrial versus ventricular pacing in sick sinus syndrome. A prospective randomized trial in 225 consecutive patients. Eur Heart J 1993; 14 (Abstr Suppl): 252.
- [89] Sutton R, Kenny RA. The natural history of sick sinus syndrome. Pace 1986; 9 (II): 1110-4.
- [90] Fisher J, Furman S, Escher D. Pacing in the sick sinus syndrome. Profile and prognosis. In: Feruglio G, eds. Cardiac pacing, electrophysiology and pacemaker technology. Padova Piccin 1982; 519-20.
- [91] Hauser R, Jones J, Edwards L, Messer J. Prognosis of patients paced for AV block or sinoatrial disease in the absence of ventricular tachycardia. Pace 1983; 6: A123.
- [92] Alt E, Völker R, Wirtzfeld A, Ulm K. Survival and follow-up after pacemaker implantation: A comparison of patients with sick sinus syndrome, complete heart block and atrial fibrillation. Pace 1985; 8: 849-55.
- [93] Alpert M, Curtis J, Sanfelippo J et al. Comparative survival following permanent ventricular and dual-chamber pacing for patients with chronic symptomatic sinus node dysfunction with and without congestive heart failure. Am Heart J 1987; 113: 958-65.
- [94] Alpert M, Curtis J, Sanfelippo J et al. Comparative survival after permanent ventricular and dual chamber pacing for patients with chronic high degree atrioventricular block with and without preexistent congestive heart failure. J Am Coll Cardiol 1986; 7: 925–32
- [95] Linde-Edelstam C, Gullberg B, Norlander R, Pehrsson SK, Rosenqvist M, Rydén L. Longevity in patients with high degree atmoventricular block paced in the atrial synchronous or the fixed rate ventricular inhibited mode. Pace 1992; 15: 304-12.
- [96] Elshot SRE, El Gamal MIH, Tielen KHJ, Van Gelder BM. Incidence of atrioventricular block and chronic atrial flutter/ fibrillation after implantation of atrial pacemakers; follow-up of more than ten years. Int J Cardiol 1993; 38: 303-8.
- [97] Bellinder G, Nordlander R, Pehrsson SK, Aström H. Atrial pacing in the management of sick sinus syndrome: Longterm observation for conduction disturbances and supraventricular tachyarrhythmias. Eur Heart J 1986; 7: 105–9.
- [98] Rosenqvist M, Arén C, Kristensson BE, Nordlander R, Schüller H. Atrial rate-responsive pacing in sinus node disease. Eur Heart J 1990; 11: 537-42.
- [99] Sgarbossa EB, Pinski SL, Castle LW, Trohamn, RG, Maloney JD. Incidence and predictors of loss of pacing in the atrium in patients with sick sinus syndrome. Pace 1992; 15 (II): 2050-4.
- [100] Rosenqvist M, Obel IWP. Atrial pacing and the risk for atrioventricular block — time to change attitude? Pace 1988; 11: 16-22.
- [101] Attuel P, Pellegrin D, Mugica J, Coumel P. DDD pacing: An effective treatment modality for recurrent atrial arrhythmias. Pace 1988; 11: 1647-54.
- [102] Denjoy I, Leclerq JF, Druelles P, Daubert C, Coumel P. Comparative efficacy of permanent atrial pacing in vagal atrial arrhythmias and in bradycardia-tachycardia syndrome. Pace 1989; 12 (II): 1236.
- [103] Leclercq C, Mabo P, Le Helloco A, Kermarrec A, Gras D, Daubert C. Hemodynamic interest of preserving a normal sequence of ventricular activation in permanent cardiac pacing. J Am Coll Cardiol 1992; 19: 150A (Abstr).
- [104] Cui C, Yang D, Fu W et al. Hemodynamic effects of ventricular contraction sequence by AV junctional pacing in patients with SSS. J Am Coll Cardiol 1992; 19. 66A (Abstr).
- [105] Yamamoto K, Kodama K, Masuyama T et al. Role of atrial contraction and synchrony of ventricular contraction in the optimisation of ventriculoarterial coupling in humans. Br Heart J 1992; 67: 361-7.

- [106] Harper GR, Pina IL, Kutalek SP. Intrinsic conduction maximizes cardiopulmonary performance in patients with dual chamber pacemakers. Pace 1991; 14 (II): 1787-91
- [107] Raichlen JS, Campbell FW, Edie RN, Josephson ME, Harken AH. The effect of the site of placement of temporary epicardial pacemakers on ventricular function in patients undergoing cardiac surgery. Circulation 1984; 70: 1-118– I-123.
- [108] Heyndrick GR, Vilaine JP, Knight DR, Vatner SF Effects of altered site of electrical activation on myocardial performance during inotropic stimulation. Circulation 1985; 71: 1010-6.
- [109] Mueller X, Sadeghi H, Kappenberger L Complications after single versus dual chamber pacemaker implantation. Pace 1990, 13: 711-4.

- [110] Goldman B, Fraser J, Morgan C. Survey of cardiac pacing in Canada. Can J Cardiol 1991; 7. 391-8.
- [111] Parsonnet V, Bernstein A, Lindsay B. Pacemakerimplantation complication rates: An analysis of some contributing factors. J Am Coll Cardiol 1989; 13: 917-21
- [112] Eagle K, Mulley A, Singer D, Schoenfeld D, Harthorne J, Thibault G Single-chamber and dual-chamber cardiac pacemakers: A formal cost comparison Ann Int Med 1986, 105: 264-71.
- [113] Ray SG, Griffith MJ, Jamieson S, Bexton RS, Gold RG. Impact of the recommendations of the British Pacing and Electrophysiology Group on pacemaker prescription and the immediate costs of pacing in the Northern Region. Br Heart J 1992; 68: 531-4.